

Following the ICRP recommendation concerning the lowering of the dose limit for the crystalline lens, LANDAUER® set up a working group on this topic in 2012: the GEDOC. The group has conducted a number of workstation studies in medical, nuclear and industrial sectors. The VISION lens dosimeter design is based on what our clients said.







### A DOSIMETER DESIGNED FOR YOU

- Worker monitoring and workstation study
- Multiple ways to wear it with or without PPE
- Adaptable to any support
- Clear, complete and nonerasable identification
- Individual dose equivalent H<sub>p</sub>(3)
- Disinfection by immersion or brushing





# VISION® monitoring your lens in discreet

Ergonomic lens dosimeter without blocking your field of vision

# >

#### SIMPLE AND PRACTICAL TO USE

#### Compact and flexible

Ergonomic, light and small, VISION dosimeter does not affect participant's activity and view. It is made from a plastic resistant to torsion. Flexible, it can be carried closer to the eye.

#### Flexible design

VISION is adapted to workstation studies or worker monitoring. It is adaptable, you can configure it for measurements behind or in front of the PPE (Personal Protective Equipment).

Offered in two sizes, VISION is adapted to all supports: sealed glasses, visors, mob caps... Self-gripping pads ensure a maximum fit.

Available on option: non-lead glasses to wear it without PPE.

The cap can be oriented from vertical to horizontal position to fit your morphology and your PPE.



#### **>** V

#### WITH OR WITHOUT PPE, VISION FITS YOUR VIEW

with PPE



Lead



Lead protection



without PP

Mob caps



Option: holding glasses (non- leaded glasses)



# A MULTITUDE OF WAYS TO WEAR IT

You receive VISION standard configuration.

Following your use (participant monitoring, workstation studies, wearing of PPE, ...) you can adapt on the clamp:

- the position of the cap, in front or behind your PPE,
- the orientation of the cap, lateral or horizontal.

Once the position and orientation defined clip the cap on the clamp.





### >

# **COMPLETE AND NONERASABLE IDENTIFICATION**

With the same color code than IPLUS® and MONORING®, VISION offers a clear identification of the wearer and period of wear.

Particles of titanium oxide are injected into the polyethylene cap. They are revealed by laser for a permanent marking. Thus, it resists brushing and all disinfection procedures.





# ways

#### **EASY TO DISINFECT**

VISION dosimeter clamp is made with Polyamide 6 and SEBS\* molding. TLD chip is sealed in a linear low density polyethylene cap. Thus, VISION is easy to clean and disinfect.

#### Maximum hygiene

Simple disinfection protocols, same as MONORING, are available on request. It can also be decontaminated easily using wipes.

We do not recommend heat sterilization whose temperature exceeds 40  $^{\circ}\text{C}$ .



#### Disinfectants suitable for use on VISION

Peracetic acid	Compatibility
Anioxide 1000	Yes
Anioxy-twin*	Yes
Endocide	Yes
Concentrated hydraseptic	Yes
NU Cidex ASP	Yes
Chlorine derivatives	Compatibility
Wip'Anios Premium* wipes	Yes
Hydro-alcohol solutions	Compatibility
Ethyl alcohol	Yes
Isopropyl alcohol	Yes
Hydrogen peroxide	Compatibility
Actanios HLD*	Yes
Bioxal M	Yes, 1hr maximum contact
Pre-disinfection products	Compatibility
Salvanios pH7	Yes
Septanios	Yes

These data come from theoretical material compatibility tables. Instructions must be validated by our your health committee.

# **MODULARITY**

VISION is designed for worker mandatory monitoring as well as workstation studies

# TECHNICAL PERFORMANCE

Types of radiation measured	Result of the VISION dosimeter
	Photons
Personal dose equivalent	$H_{\triangleright}(3)$
Dose range	from 0.10 mSv to 10 Sv
Minimum reporting value	0.1 mSv
Linearity response	from 0.10 mSv to 10 Sv Standard deviation <9 %
Energy response (average energy)	from 24 keV to 1.25 MeV
Angular response	± 60° from 24 keV to 1.25 MeV
Response to other types of radiation	Insensitive to neutrons



# **ENVIRONMENTAL RESISTANCE CHARACTERISTICS**

Operating and storage temperature	No effect on detector sensitivity up to 40 °C
Hygrometry	No effect on detector up to 90 % humidity
Exposure to light	No effect on detector

# GENERAL CHARACTERISTICS

Manufacturer Dosimeter designed and manufactured in France for LANDAUER

Detector manufactured according to LANDAUER's own specifications

Types of radiation

measured Photons (X- and gamma rays)

Detector Single element, one TLD chip

Chip material LiF:Mg,Ti (lithium fluoride doped with magnesium and titanium)

Clamp material Polyamide 6 (SEBS molding)

Cap material Polyethylene

Mechanical<br/>propertiesWeightM: 720 mgL: 770 mg

L: 770 mg

Two sizes M: 4.6 mm

L: 7.6 mm



### **MEASUREMENT METHOD**

VISION dosimeter is composed of a polyamide clamp on which is clipped a cap. It contains a lithium fluoride TLD chip. This detector is placed inside a cavity on the cap which is sealed by ultrasound. The information is marked on the cap by laser, making it nonerasable.

The TLD (ThermoLuminescent Dosimeter) works on the basis of detector heating. When the chip is heated, visible light is emitted in proportion to the exposure of ionising radiation.





#### **COMPLIANCE WITH STANDARDS**

EN 62387-1:2012 - Passive integrating dosimetry systems for personal and environmental monitoring of photon and beta radiation - Radiation protection instrumentation

# >

#### **QUALIFICATIONS OF OUR LABORATORY**

- Participation in national and international inter-comparisons
- Dosimeter characterisation carried out by an independent reference laboratory: Henri Becquerel French National Laboratory (LNHB) - CEA

